Ref. Certif. No.

BE-57845

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	6 channels DC Energy Meter
Name and address of the applicant	Acrel Co., Ltd. No.253, Yulv Road, Jiading, Shanghai, China
Name and address of the manufacturer	Acrel Co., Ltd. No.253, Yulv Road, Jiading, Shanghai, China
Name and address of the factory Note: *When more than one factory, please report on page 2*	Jiangsu Acrel Electrical Manufacturing Co., Ltd. No.5, Dongmeng Road, Nanzha, Jiangyin, Jiangsu, China ☐ Additional Information on page 2
Ratings and principal characteristics	Input: DC 0-48V, DC 5V, auxiliary power: -48V~-60VDC Power≤2W 32.768kHz
Trademark (if any)	Acrel
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	AMC16-DETT, AMC16L-DETT
Additional information (if necessary may also be reported on page 2)	- ☐ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016 IEC 61010-2-030:2023 National Differences: EU Group Differences
As shown in the Test Report Ref. No. which forms part of this Certificate	SHES260300435501

This CB Test Certificate is issued by the National Certification Body

SGS Belgium NV - Division SGS CEBEC
Riverside Business Park
Bld Internationalelaan 55, Building A
B-1070 Brussels, Belgium

Date: 2026-06-11

Signature: Mark Lohmann

This certificate is issued by the company under its General Conditions for Certification Services accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx
Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this certificate. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.